



SIDING PREPARATION  
TYPICAL DETAIL  
SCALE: 3/8=1"

NOTE:  
LEAD PANEL SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET, FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION-QLQ201. HUTCH SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANSI N43.2 AND DOE ORDER 5480.11.

- DIMENSION TO BE DETERMINED BY FIELD SURVEY

ITEM	QUANTITY	UNIT	DESCRIPTION	MATERIAL / SPEC	REV
PARTS LIST					
ARGONNE NATIONAL LABORATORY					
A2170600					
ADVANCED PHOTON SOURCE					
HO ID FOE ROOF # 24					
1/2" x 2.24M x .50M					
P41050903-980304-00					